

Expanded readout



Use the EXP readout mode when you wish to see the results with one additional decimal place of accuracy. The EXP Key toggles the readout back and forth between standard to expanded display. The table below shows the result of using the EXP readout mode for each of the six parameters.

Table 1. Accuracy of expanded readout




| Parameter | Range of measurement | Accuracy | |
|-----------|----------------------|------------------|------------------|
| | | Standard readout | Expanded readout |
| pH | 0-14 pH | 0.1 pH | 0.01 pH |
| COND | 0-1 mS/cm | 0.01 mS/cm | 0.001 mS/cm |
| | 1-10 mS/cm | 0.1 mS/cm | 0.01 mS/cm |
| | 10-100 mS/cm | 1 mS/cm | 0.1 mS/cm |
| TURB | 0-800 NTU | 10 NTU | 1 NTU |
| DO | 0-19.9 mg/l | 0.1 mg/l | 0.01 mg/l |
| TEMP | 0-50°C | 1°C | 0.1°C |
| SAL | 0-4% | 0.1% | 0.01% |


Note that the salinity parameter is the only value not measured directly with its own sensor. The U-10 obtains salinity by converting the conductivity value. If large amounts of conductive ions other than salt-water components are present in the sample, an error may occur. Be cautious when interpreting the salinity results.



Measuring fresh water or salt water?

The U-10 can be set to the salinity for either fresh water or salt water when measuring DO. This is done by using the S.SET Sub-Mode.

Measuring fresh water

-  1. First, use the MODE Key to put the U-10 in the MAINT mode. Keep pressing the MODE Key to toggle the lower cursor to the S.SET Sub-Mode.
-  2. Once you are in the S.SET Sub-Mode, use the UP/DOWN Keys to select the salinity value. For fresh water, set the salinity to 0.0%.
- 



-  3. Finally, press the ENT Key to complete the salinity setting while in the S.SET Sub-Mode.
-  4. When the salinity setting has been made, switch back to the MEAS mode by pressing the the MODE Key.

Measuring salt water




1. First, use the MODE Key to put the U-10 in the MAINT mode. Keep pressing the MODE Key to toggle the lower cursor to the S.SET Sub-Mode.



2. For salt water, set it to *A* i.e., for auto-salinity.



The *A* setting should be sufficient for measurements of normal sea water with a salinity value close to 3.3%. For sea water of an unusual salinity, however, and where the value is otherwise known, you may wish set the value manually to any salinity within the range of 0.0%-4.0%. (You may also possibly want to use a manual setting if, for example, the COND sensor is malfunctioning but it is still desirable to take readings of the other parameters.)



3. Finally, press the ENT Key to complete the salinity setting while in the S.SET Sub-Mode.



4. When the salinity setting has been made, switch back to the MEAS mode by pressing the the MODE Key.

After measurement: Cleaning and storing the U-10



1. Turn OFF the power.
2. Wash the probe thoroughly with tap water. Be sure to flush off all of sample solution from the probe.

Storing the U-10 for brief periods, *i.e.*, about 1 week or less:

Fill the calibration beaker with tap water and fit the probe over it.

For longer storage

The pH sensor must always be kept moist. Fill the small rubber cap with water and use it to cover the pH sensor.

The KCl internal solution in the reference sensor may seep out over time. Place vinyl tape around the O-ring portion to prevent this.

If you are going to store the U-10 for a prolonged period without using it, remove the battery from the main unit.

3

Section

Calibrating the U-10

The U-10 Water Checker may be calibrated either manually or automatically. The 4-parameter auto-calibration procedure is quite handy and should be sufficient for most measurement operations.

Manual calibration for each of the four parameters is more accurate but, of course, also more time-consuming. This method should be used for more precise measurement. The manual calibration procedure is explained below in detail, following the description of the auto-calibration procedure.


The auto-calibration procedure is extremely simple. The U-10 Water Checker uses just a single solution to do a simultaneous calibration of four parameters: *pH*, *COND*, *TURB*, and *DO*. Your U-10 comes with a bottle of standard phthalate pH solution and a calibration beaker for this purpose.

| | |
|--|----|
| Auto-calibration procedure | 20 |
| Manual (2-point) calibration procedures | 23 |
| pH Calibration | 24 |
| 1.Zero calibration | 24 |
| 2.Span calibration | 25 |
| COND Calibration | 26 |
| 1.Zero calibration | 28 |
| 2.Span calibration | 29 |
| TURB Calibration | 30 |
| 1.Zero calibration | 31 |
| 2.Span calibration | 31 |
| DO Calibration | 32 |
| 1.Zero calibration | 33 |
| 2.Span calibration | 33 |


Auto-calibration procedure

Fill the calibration beaker to about 2/3 with the standard solution. Note the line on the beaker.

Fit the probe over the beaker, as illustrated. Note that the beaker is specially shaped to prevent the DO sensor from being immersed in the standard solution. This is because the DO auto-calibration is done using atmospheric air.




With the power on, press the MODE Key to put the unit into the MAINT mode. The lower cursor should be on the AUTO Sub-Mode; if it is not, use the MODE Key to move the lower cursor to AUTO.




With the lower cursor on AUTO, press the ENT Key. The readout will show *Err*. Wait a moment, and the upper cursor will gradually move across the four auto-calibration parameters one-by-one: *pH*, *COND*, *TURB*, and *DO*. When the calibration is complete, the readout will briefly show *End* and then will switch to the MEAS mode.


The upper cursor will blink while the auto-calibration is being made. When the auto-calibration has stabilized, the upper cursor will stop blinking.




First, pH is being auto-calibrated




Then, COND is being auto-calibrated




Next, TURB is being auto-calibrated



Finally, DO is being auto-calibrated



Auto-calibration now ends




And the readout switches to the MEAS mode

Note: If you wish to abort the auto-calibration for any reason, press the CLR Key. The parameters auto-calibrated so far will be stored in memory.

Auto-calibration error


After the DO auto-calibration, if the unit does not switch to the MEAS mode as it should, and the readout shows either *E-3* or *E-4*, an auto-calibration error has occurred. Parameters will blink where an error occurred.



pH auto-calibration error

CLR

If this happens, re-do the auto-calibration. First, press the CLR Key to cancel the error code.



ENT

Then press the ENT Key to re-start the auto-calibration. Restart the auto-calibration beginning again with pH.

Manual (2-point) calibration procedures

For normal measurements, the 4-parameter auto-calibration described above is sufficiently accurate. However, you may wish to do a parameter-by-parameter, 2-point manual calibration of one or more of the four parameters. This is recommended either for high-accuracy measurements, especially when using the expanded readout mode. It is necessary if a new probe is being used for the *first time*.

Parameters to be calibrated manually.

| | |
|------|-----------------------|
| pH | • Zero (see page 24.) |
| | • Span (see page 25.) |
| COND | • Zero (see page 28.) |
| | • Span (see page 29.) |
| TURB | • Zero (see page 31.) |
| | • Span (see page 31.) |
| DO | • Zero (see page 32.) |
| | • Span (see page 33.) |

Parameters not to be calibrated.

Sample temperature
Salinity

pH calibration

pH calibration on the U-10 is done using two commercially-available standard solutions of different pH values, one for the zero calibration, the other for the span calibration. Note that the temperature characteristics of the various standard solutions that are available may differ; therefore, before using these two solutions to make the pH calibration, carefully measure the temperature and determine the temperature characteristics of each.

Preparation

Wash the probe 2-3 times, using de-ionized or distilled water. Place it in a beaker of each standard solution.

1. Zero calibration

Use a pH7 standard solution for the zero calibration.

Operation



1. With the power on, press the MODE Key to put the unit into the MAINT mode.





2. Press the MODE Key again to move the lower cursor to ZERO.



3. Use the SELECT Key to move the upper cursor to pH.



4. When the readout has stabilized, use the UP/DOWN Keys to select the value of the pH 7 standard solution at the temperature of the sample. Refer to Table 2 for pH values of standard solutions at various temperatures.



5. Press the ENT Key to complete the zero calibration for pH.

2. Span calibration

Use either a pH4 or a pH9(10) standard solution for the span calibration.

Operation



1. Use the MODE Key to move the lower cursor to SPAN.



2. As in Step 4. above in zero calibration, when the readout has stabilized, use the UP/DOWN Keys to select the value of the standard solution (i.e., either pH4 or pH9) at the temperature of the sample.



Again, refer to Table 2 for pH values of standard solutions at various temperatures.



3. Press the ENT Key to complete the span calibration for pH.




Table 2 pH values of standard solutions at various temperatures*

| Temperature °C / °F | pH2 ^a | pH4 ^b | pH7 ^c | pH9 ^d | pH10 ^e | pH12 ^f |
|------------------------|------------------|------------------|------------------|------------------|-------------------|-------------------|
| 0 / 32 | 1.67 | 4.01 | 6.98 | 9.46 | 10.32 | 13.43 |
| 5 / 41 | 1.67 | 4.01 | 6.95 | 9.39 | 10.25 | 13.21 |
| 10 / 50 | 1.67 | 4.00 | 6.92 | 9.33 | 10.18 | 13.00 |
| 15 / 59 | 1.67 | 4.00 | 6.90 | 9.27 | 10.12 | 12.81 |
| 20 / 68 | 1.68 | 4.00 | 6.88 | 9.22 | 10.06 | 12.63 |
| 25 / 77 | 1.68 | 4.01 | 6.86 | 9.18 | 10.01 | 12.45 |
| 30 / 86 | 1.69 | 4.01 | 6.85 | 9.14 | 9.97 | 12.30 |
| 35 / 95 | 1.69 | 4.02 | 6.84 | 9.10 | 9.93 | 12.14 |
| 40 / 104 | 1.70 | 4.03 | 6.84 | 9.07 | 9.89 | 11.99 |
| 45 / 113 | 1.70 | 4.04 | 6.83 | 9.04 | 9.86 | 11.84 |
| | 1.71 | 4.06 | 6.83 | 9.01 | 9.83 | 11.70 |

a : oxalate, b : phthalate, c : neutral phosphate, d : borax,
e : carbonate, f : Sat.calcium hydroxide solution

* These pH values are for Japanese standard solutions. Should you prefer to use different standard solutions, be sure to make the proper adjustments in calibration.

COND calibration

The U-10 can measure conductivity in the range of 0-100 mS/cm. Depending on the sample concentration, however, the U-10 automatically selects the proper range out of its three possible ranges of 0-1 mS/cm, 1-10 mS/cm, and 10-100 mS/cm.

Therefore, if you are doing a manual calibration for COND, this must be done for each of the three ranges. However, since the zero point is common for all three ranges, only the three one-point span calibrations need be done separately.

Preparing the standard solution for COND span calibration

This solution uses a potassium chloride as a reagent. For greater accuracy, the solution should be freshly prepared each time. If it is unavoidable to use a stored solution, be sure to keep it tightly capped in a polyethylene or hard glass bottle. The shelf life of this solution is six months. Date-stamp the bottle for reference. Never use a KCl standard solution that has been stored for more than six months: the calibration accuracy may be adversely affected.

Use potassium chloride powder of the best quality commercially available. Dry the powder for two hours at 105°C, and cool it down, in a desiccator. Weigh out an appropriate amount of dried and cooled potassium chloride powder according to the table below. Make the potassium chloride standard solution as shown.

Table 3 Making the potassium chloride standard solution

| KCl standard solution | KCl weight g | Conductivity* mS/cm | Range to be calibrated mS/cm |
|-----------------------|--------------|---------------------|------------------------------|
| 0.005N | 0.373 | 0.718 | 0-1 |
| 0.05N | 3.73 | 6.67 | 1-10 |
| 0.5N | 37.28 | 58.7 | 10-100 |

* Value at the temperature, 25°C

To prepare the standard solution, use a 1-liter volumetric flask. First, dissolve the KCl in a small amount of de-ionized or distilled water. Then fill the flask with de-ionized or distilled water up to the 1-liter line. Finally, shake the solution to mix it thoroughly.

1. Zero calibration

This calibration is carried out in atmospheric air; no solution is needed.

Preparation

Wash the probe 2-3 times, using de-ionized or distilled water. Shake the probe to remove any water droplets from the COND sensor. Then allow it to dry by exposing it to fresh air.

Operation





1. Use the MODE Key to move the lower cursor to ZERO.



2. Use the SELECT Key to move the upper cursor to COND.



3. Use the UP/DOWN Keys to set the readout to zero.



4. Press the ENT Key. This completes the zero calibration for COND.

2. Span calibration

This procedure uses a standard solution of potassium chloride. For best results, a fresh batch of the solution should be prepared each time. See page 27 for details.

Preparation

Wash the probe 2-3 times using de-ionized or distilled water. Following this, wash it 2-3 times in the KCl standard solution you have prepared. Then place the probe in a beaker of the KCl solution maintained at a temperature of $25 \pm 5^\circ\text{C}$.

Operation



1. Use the MODE Key to move the lower cursor to SPAN.



2. After the readout stabilizes, as you did for the pH calibration, use the UP/DOWN Keys to select set the value of the KCl standard solution, referring to the KCl table.



3. Press the ENT Key to complete the span calibration for this COND range.
4. Repeat this procedure for the three ranges, using each of three values of KCl standard solutions.

TURB calibration

Use good-quality de-ionized water, which may be considered as having a turbidity of zero. If that is not readily available, distilled water may be used instead. When doing the turbidity zero calibration, it is particularly crucial that you clean the probe thoroughly. Never use a dirty probe; otherwise the calibration will be unreliable.

Preparing the standard solution for TURB span calibration

1. Weigh out 5.0 g of hydrazine sulfate.
2. Dissolve this in 400 ml of de-ionized or distilled water.
3. Then weigh out 50 g of hexamethylenetetramine, and dissolve it in 400 ml of de-ionized or distilled water.
4. Mix these two solutions, add enough de-ionized or distilled water to make 1,000 ml, and stir the mixed solution thoroughly.
5. Allow this solution to stand for 24 hours at a temperature of $25 \pm 3^{\circ}\text{C}$.

The turbidity of this solution is equivalent to 4000 NTUs. The shelf-life of this solution is six months; i.e., this 4,000-NTU value will remain accurate for a maximum of six months.

Each time you carry out this calibration, it is necessary to dilute the 4,000-NTU standard solution to prepare an 800-NTU standard solution for calibration. To do this, measure out 50 ml of the 4,000-NTU solution into a 250-ml measuring flask.

It is recommended that you use a rubber pipette aspirator for this. Then add de-ionized or distilled water up to the 250-ml line.


The standard solution used here for the turbidity calibration will precipitate easily. Therefore, be sure to stir the solution thoroughly before use.

1. Zero calibration

Preparation

Wash the probe thoroughly 2-3 times using de-ionized or distilled water. Shake off excess water droplets, and then place it in a beaker of de-ionized or distilled water.

Operation




1. Use the MODE Key to move the lower cursor to ZERO.
2. Use the SELECT Key to move the upper cursor to TURB.
3. After the readout has stabilized, set it to 0.0, using the UP/DOWN Keys.
4. Press the ENT Key to complete the zero calibration for TURB.

2. Span calibration

Preparation

Wash the probe thoroughly, using de-ionized or distilled water. Shake off excess water droplets. Then place it in a beaker of the 800-NTU solution you have prepared for this purpose.

Operation



1. Stir this 800-NTU span standard solution thoroughly.
2. Use the MODE Key to move the lower cursor to SPAN.
3. After readout has stabilized, i.e., about 60 to 90 seconds, set the readout to "800" NTU, which is the value for this standard solution.
4. Press the ENT Key to complete the span calibration for TURB.

DO calibration

Unlike the other calibration procedures, the solution for the DO calibration cannot be stored for use; because the amount of dissolved oxygen in the solution is crucial, a fresh batch must be prepared each time, just before it is used in the DO calibration.


1. Zero calibration

Use a solution of sodium sulfite dissolved in either de-ionized water or tap water.

Preparation

1. Add about 50g of sodium sulfite to 1,000 ml of water (either de-ionized water or tap water will do). Stir this mixture to dissolve.
2. Wash the probe 2-3 times in tap water, and place it in the zero standard solution.

Operation



1. Use the MODE Key to move the lower cursor to ZERO.
2. Use the SELECT Key to move the upper cursor to DO.
3. After the readout has stabilized, set it to 0.0, using the UP/DOWN Keys.
4. Press the ENT Key. This completes the zero calibration for DO.

2. Span calibration

Use either de-ionized water or tap water that has been saturated with oxygen in air.

Preparation

1. Put 1 or 2 liters of water in a container (either de-ionized water or tap water will do). Use an air pump to bubble air through the solution until it is oxygen-saturated.
2. Wash the probe 2-3 times in tap water, and put it in the span calibration solution.

Operation

1. First, be sure the U-10 is set for fresh water readings. To do this, set the S.SET Sub-Mode to 0.0%.
2. Then, use the MODE Key to move the lower cursor to SPAN.
3. After the readout has stabilized, while slowly moving the probe up and down in the solution, set the readout value to the appropriate DO value for the temperature of this solution. For DO values at various temperatures, refer to Table 4.
4. Press the ENT Key to complete the span calibration for DO.




Table 4 Amounts of saturated dissolved oxygen in water at various temperatures, salinity = 0.0%

| Temperature | DO | Temperature | DO |
|-------------|------------|-------------|-----------|
| 0 °C | 14.16 mg/l | 21 °C | 8.68 mg/l |
| 1 | 13.77 | 22 | 8.53 |
| 2 | 13.40 | 23 | 8.39 |
| 3 | 13.04 | 24 | 8.25 |
| 4 | 12.70 | 25 | 8.11 |
| 5 | 12.37 | 26 | 7.99 |
| 6 | 12.06 | 27 | 7.87 |
| 7 | 11.75 | 28 | 7.75 |
| 8 | 11.47 | 29 | 7.64 |
| 9 | 11.19 | 30 | 7.53 |
| 10 | 10.92 | 31 | 7.42 |
| 11 | 10.67 | 32 | 7.32 |
| 12 | 10.43 | 33 | 7.22 |
| 13 | 10.20 | 34 | 7.13 |
| 14 | 9.97 | 35 | 7.04 |
| 15 | 9.76 | 36 | 6.94 |
| 16 | 9.56 | 37 | 6.86 |
| 17 | 9.37 | 38 | 6.76 |
| 18 | 9.18 | 39 | 6.68 |
| 19 | 9.01 | 40 | 6.59 |
| 20 | 8.84 | | |