### **REVOLUTIONARY 12 VOIT DC PUMPS** FOR GROUNDWATER PURGING OR LOW-FLOW SAMPLING FOR 2" OR LARGER DIAMETER WELLS





# **OPERATING INSTRUCTIONS**



\* Depth to Water

Supplied with 90 feet of heavy duty 12 gauge wire

Not Compatilbe with any ProActive Controllers or Boosters

The Proactive Super Twister, Cyclone, Mini Typhoon, Typhoon, Tempest, Mini-Monsoon, Mega-Typhoon, Tsunami, Monsoon, Hurricane, Mega-Monsoon, Stainless Steel Mega-Typhoon and Super Twister are Patent Pending

## **OPERATING INSTRUCTIONS FOR THE SS-MEGA-TYPHOON**

#### Simple to Use and Extrememly Powerful

The SS MEGA-TYPHOON is capable of pumping up to 80 feet from ground level to water level. Simply connect the pump connector to the Low Flow with Power Booster I "LCD" Controller, then to a 12 volt battery, energizing the pump. Sound Simple? We think so.

#### "60 Second Change Out Motor Design!"

The SS MEGA-TYPHOON is practically maintenance free! The only part that may need to be serviced is the replaceable motor module. Should the existing replaceable motor module fail in the field, a new replaceable motor module can quickly be installed within 60 seconds by a simple quarter turn twist. Once installed, your pump is essentially new and will have a life span identical to a new pump.

The main wear out parts consisting of the motor and seal all have been incorporated into a single replaceable motor module.

#### The SS-Mega-Tuphoon is only 1.82" OD and constructed of three high polished stainless steel parts:



#### Changing Out The Motor Module – Easy as 1...2...3...

- 1. Unscrew the outside pump housing. If needed use the included ProActive wrench (Fig. D).
- 2. Ouarter twist off replaceable motor module (Fig. B).
- Insert the two new o-rings included, onto the motor housing top connector (Fig. A).
- 3. Slightly insert the replaceable motor module onto the motor housing top connector. Be gentle, not to rip the o-rings when replacing them back onto the motor housing top connector!! Water will leak into the replaceable motor module if the o-rings are cut or torn... shortening motor life!!
- 4. Line up the 3 cut-away grooves on the replaceable motor module with the 3 location dowels on the motor housing top connector. Then slowly twist the replaceable motor module until tightened. Once again, be gentle not to rip the o-rings when inserting the replaceable motor module onto the motor housing top connector!! Water will leak into the replaceable motor module if the o-rings are cut or torn... shortening motor life!!
- 5. Tighten the outside pump housing. If needed use the included ProActive wrench (Fig. D).
- 6. Turn on pump dry (not in water) to test pump. Leave running for 10 seconds. If the motor turns on, then the pump is ready for use.
- 7. If the motor fails to turn on, please repeat steps 1 through 6.

#### Using the Low Flow With Power Booster I "LCD" Controller

This controller is very simple to use and requires minimal maintenance. The flow control dial is used to adjust the flow of water being discharged. From very low flow to high output please always use the LCD display as a guide to determine controller voltage output; Which ultimately controls the pump's RPM'S (revolutions per minute or speed) As you increase the dial (turn clockwise) so does the voltage. This will enable the pump motor to spin faster, therefore, increasing the flow of water being discharged. To slow the flow of water, turn the dial counter-clockwise. Please note, if the pump should ever "pulsate" indicates that your battery is critically low (below 11 volts). In short, the controller is telling you to charge your battery.

(1) Turn dial on controller all the way off. (2) Connect controller to 12 volt source. (3) Connect pump to controller. (4) Lower pump into well with 3/8" ID Tubing. (5) When pump

reaches designated depth in water, slowly increase controller voltage until desired water flow is achieved. (6) To obtain the lowest possible flow rate, make sure the dial on the controller is set to the Sampling Valve lowest voltage reading that makes for your "best control flow". (7) To achieve a very low flow rate, simply connect the PVC Low Flow Sampling Valve (Fig. F) to the end of the discharge hose. Then use the dial on the valve to adjust a perfect low flow sample. This valve will allow the user to obtain 10 ML Per Minute

or lower flow rates. NOTE: Do not use the control flow valve when the discharge rate is greater than 100 MI Per Minute. This will cause premature motor damage.

#### If your pump fails to work-Please follow the trouble shooting quide below.

- 1. Battery is very low. Check battery level. Must be at least 11 volts minimum.
- Blown fuse in controller-needs to be replaced. Replaceable Motor Module needs to be
- exchanged with a new one.
- Make sure all connectors are plugged into correct outlets and make sure controller dial is
- turned all the way up (turned clock wise all the
- way until it stops).5. If the pump should ever "pulsate" indicates that your battery is critically low (below 11 volts) in short, the controller is telling you to charge your battery.

#### Warranty Information

Proactive will warranty the SS MEGA-TYPHOON to be free from defects in material and workmanship affecting its use for a period of 60 days following the date of purchase.

#### **Exclusions and Conditions**

The use of any pump with replacement parts or accessories, which are not manufactured by ProActive and which affect product safety or performance, shall render this warranty and all other warranties, whether express or implied, null and void. Should any failure to conform to this warranty appear within the term of initial sale, the manufacturer shall upon notification thereof and substantiation that the pump has been maintained, correct such defects by suitable repair or replacement at its own expense, purchaser being responsible for shipping and handling charges only.

The Proactive warranty does not include shipping or an obligation to provide another unit while the original SS MEGA-TYPHOON pump is being repaired or replaced.

The Proactive warranty does not apply to any SS MEGA-TYPHOON pump used as rental equipment or contaminated by materials harmful to such products.

The Proactive warranty does not apply to parts failure due to neglect in cleaning or unauthorized servicing the Proactive SS MEGA-TYPHOON pump nor does it cover failure of a part caused by misuse or inappropriate use of the SS MEGA-TYPHOON submersible pump. ProActive will not be held liable for any "down time" due to a defective SS MEGA-TYPHOON Pump.

#### Returning the SS MEGA-TYPHOON Submersible Pump to Proactive

Proactive requires that all customers notify Proactive prior to returning a defective SS MEGA-TYPHOON submersible pump under the Proactive warranty program. The SS MEGA-TYPHOON will only be accepted if it has been thoroughly cleaned and decontaminated. In addition, ProActive's SS MEGA-TYPHOON must be properly packaged and accompanied by a receipt of purchase.

Please contact Proactive to obtain a RMA #

#### **IMPORTANT NOTES:**

"DO NOT TOUCH INSIDE OF PUMP IF RUNNING CONTINOUSLY FOR OVER 1 MINUTE. REPLACEABLE MOTOR MODULE GETS EXTREMLY HOT WHEN IN USE." DO NOT TURN PUMP ON IF THE OUTSIDE PUMP HOUSING IS NOT ATTACHED TO THE MOTOR HOUSING TOP CONNECTOR. HAND INJURY MAY OCCUR FROM FAST SPINNING IMPELLAR!!

Please make photo-copy of this manual when returning pump. Please return pump to:



1767 Lakewood Ranch Boulevard, Suite #113 • Bradenton, FL 34211 Phone: 1.941.322.8504 • Fax: 1.941.322.8519

Web Address: www.gopronow.biz • Email Address: proactive@gopronow.biz



"LCD" CONTROLLER

Fig. F - The

**PVC** Low Flow

POWER BOOSTER I